



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Hydrocarbon Gas Vapor Meter
 Positive Displacement Diaphragm Meters
 Model: NG4
 Capacity: 211 cubic feet per hour (CFH) @ 7.0-inch water column
 Maximum Operating Pressure: 7.25 psig

Submitted By:


Norgas Metering Technologies, Inc.
 473 Commercial Drive
 Fairfield, OH45014
 Tel: 513-942-3600 ext. #221
 Fax: 513-942-0096
 Contact: Michael Dick
 Email: miked@norgas.com
 Web site: www.norgas.com


Standard Features and Options

- Aluminum Body
- Corrosion Resistant Internal Components and External Epoxy Coating
- Buna-N Diaphragm material
- Non-temperature Compensating
- Cubic Feet Registration
- 8-Digit Analog Register
- Top Arrow Indicating Flow Direction (Left to Right)

Temperature Range: -4 °F to +122 °F (-20 °C to 50 °C)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.


 Stephen Benjamin
 Chairman, NCWM, Inc.


 Kurt Floren
 Committee Chair, National Type Evaluation Program Committee
 Issued: September 13, 2012

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Norgas Metering Technologies, Inc.
Hydrocarbon Gas Vapor Meter / NG4

Application: Approved for commercial measurement of hydrocarbon gases in a legal sub-metering installation.

Identification: The NG4 will have an identification plate glued to the front of the meter. It will have the manufacturer's name, base capacity, the NTEP CC number, and the manufacturing serial number. Other information is also on the register face.

Sealing: Two lead/plastic security plugs are used to seal the index screws. A wire security seal through the index and pulse output screw is used to seal the index and pulse outputs from being able to be removed. Ten security screws are used to secure the top and bottom meter casing along with one front and one back lead wire seal.

NOTE: The index and/or any fasteners **MUST NOT** be removed otherwise this will void the factory warranty.

Test Conditions: Two NG4 non- temperature compensating meters were submitted for evaluation. Accuracy tests were performed at various flow rates using a five cubic foot bell prover. Also a low flame test was performed. The accuracy tests were repeated after over 170 000 cubic feet of air was passed through the devices.

Evaluated By:J. Roach (CA)

Type Evaluation Criteria Used:*NIST Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices, 2012 Edition. NCWM Publication 14 Measuring Devices, 2012 Edition.*

Conclusion:The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By:J. Truex (NCWM)

Example(s) of Device:

